




CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE MUTHANENDHAL, SIVAGANGAI.		CERTIFICATE NO	TICS/2023/SG/MNL/05		
		EQUIPMENT NAME	MICROPIPETTE		
		MANUFACTURE	THERMO SCIENTIFIC		
		MODEL	100-1000µL		
CALIBRATION DATE	22/09/2023	SERIAL NO	RW02663		
CALIBRATION DUE DATE	21/09/2024	INSTALLED DEPARTMENT	LAB		
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT			Temperature 25°C Humidity RH 55%		
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.02	± 1	PASS
		1000	1000.05	± 1	PASS
* The above mentioned instrument has been tested using standards manufactures recommended protocols. using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : 		Authorized Sign With Seal 