

CALIBRATION / TESTING REPORT				
<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE MUTHANENDHAL, SIVAGANGAI.		<b>CERTIFICATE NO</b>	TICS/2023/SG/MNL/002	
		<b>EQUIPMENT NAME</b>	BIO CHEMISTRY ANALYZER	
		<b>MANUFACTURE</b>	ROBONIK	
		<b>MODEL</b>	PRIETEST TOUCH	
<b>CALIBRATION DATE</b>	22/09/2023	<b>ASSET NO</b>	ATCD0550321RBK	
<b>CALIBRATION DUE DATE</b>	21/09/2024	<b>INSTALLED DEPARTMENT</b>	LAB	
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No				
<b>TEST EQUIPMENT</b>	<b>MANUFACTURE</b>	<b>MODEL</b>	<b>SERIAL NUMBER</b>	
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011	
<b>PARAMETER</b>	<b>UNITS</b>	<b>VALUE</b>	<b>LIMITS</b>	<b>STATUS</b>
Main Voltage L1-L2	V	223.2	230±10	PASS
Main Voltage L1-G(PE)	V	220.1	230±10	PASS
Main Voltage L2-G(PE)	V	0.92	<5	PASS
Load Current	A	0.614	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.14	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS
Note : Test Standard: IEC 62353 Defibrillator paddles 100 μA. per IEC 60601-2-4				
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.				
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE				
Calibrated By : 			Authorized Sign With Seal 	

No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.

☎ 04633-210998 ☎ +91 72000 04781, +91 95001 13759

✉ transindiacalibration@gmail.com 🌐 www.transindiacalibration.com

CIN: U52100TN2020PTC139842



**CALIBRATION / TESTING REPORT**

<b>CUSTOMER DETAILS :</b>		<b>CERTIFICATE NO</b>	TIC S/2023/SG/MNL/002
GOVT.PRIMARY HEALTH CENTRE		<b>EQUIPMENT NAME</b>	BIO CHEMISTRY ANALYZER
MUTHANENDHAL,		<b>MANUFACTURER</b>	ROBONIK
SIVAGANGAL		<b>MODEL</b>	PRIE TEST TOUCH
<b>CALIBRATION DATE</b>	22/09/2023	<b>ASSET NO</b>	ATC D055032TRBK
<b>CALIBRATION DUE DATE</b>	21/09/2024	<b>INSTALLED DEPARTMENT</b>	LAB

- 1. General maintenance required  Yes  No
- 2. Any Physical damage observed  Yes  No
- 3. Spare Replacement (if required)  Yes  No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
ELECTRICAL SAFETY ANALYZER	DAIFENG	V Pad mini	IPM 18067911

PARAMETER	UNITS	VALUE	LIMITS	STATUS
Main Voltage L1-L2	V	223.2	230±10	PASS
Main Voltage L1-G(PE)	V	220.1	230±10	PASS
Main Voltage L2-G(PE)	V	0.92	<5	PASS
Load Current	A	0.614	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.14	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF - < 50μA*	PASS

Note : Test Standard: IEC 62353  
Defibrillator paddles 100 μA, per IEC 60601-2-4

\*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE       CORRECTIVE MAINTENANCE REQUIRED       REMOVED FROM USE

Calibrated By :

*Roz*

Authorized Sign With Seal





CALIBRATION / TESTING REPORT				
<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE MUTHANENDHAL, SIVAGANGAL.		CERTIFICATE NO	TICS/2023/SG/MNL/002	
		EQUIPMENT NAME	BIO CHEMISTRY ANALYZER	
		MANUFACTURE	ROBONIK	
		MODEL	PRIETEST TOUCH	
CALIBRATION DATE	22/09/2023	ASSET NO	ATCD0550321RBK	
CALIBRATION DUE DATE	21/09/2024	INSTALLED DEPARTMENT	LAB	
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No				
<b>TEST EQUIPMENT</b>	<b>MANUFACTURE</b>	<b>MODEL</b>	<b>SERIAL NUMBER</b>	
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011	
<b>PARAMETER</b>	<b>UNITS</b>	<b>VALUE</b>	<b>LIMITS</b>	<b>STATUS</b>
Main Voltage LI-L2	V	223.2	230±10	PASS
Main Voltage LI-G(PE)	V	220.1	230±10	PASS
Main Voltage L2-G(PE)	V	0.92	<5	PASS
Load Current	A	0.614	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.14	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS
Note : Test Standard: IEC 62353 Defibrillator paddles 100 μA, per IEC 60601-2-4				
* The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.				
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE				
Calibrated By :		Authorized Sign With Seal		
				
				