



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name	Govt. PHC, Vanavasi	Lab No	12597
Constituent Group	Chemistry III	Date of Result Entered :	14/09/2023
PT item	Lyophilized human serum based	Date of Report Published :	05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	68	312.35	4.82	15.05	321 mg/dL	0.57	3.65
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	69	1.31	8.62	0.11	1.4 mg/dL	0.80	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	151	113.39	4.43	5.02	143 mg/dL	5.9	0.82

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Vanavasi
VANAVASI
Vanavasi
TAMILNADU636457

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****