



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13980		Page No: 1 of 1
Customer Name & Address		
GOVERNMENT URBAN PRIMARY HEALTH CENTRE, PANJUPETTAI-631502,KANCHIPURAM DISTRICT.	SRF No.	SRF/23/00848-0001
	SRF Date	10-10-2023
	Date of Receipt	10-10-2023
	Date of Calibration	10-10-2023
	Due Date for Calibration	09-10-2024
	Issue Date	11-10-2023

Details of Unit Under Calibration			
Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW09339	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used						
SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319											
Result of Calibration in µl											
SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10.0	9.94	9.97	9.94	9.96	9.93	9.95	-0.05	0.02	0.47	
		9.98	9.95	9.92	9.97	9.96					
2	50.0	49.85	49.86	49.87	49.82	49.83	49.85	-0.15	0.02	0.47	
		49.87	49.85	49.82	49.89	49.87					
3	100.0	99.94	99.97	99.96	99.95	99.98	99.95	-0.05	0.02	0.47	
		99.93	99.95	99.97	99.94	99.95					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming to Grade 3 as specified in ISO 3696.

Calibrated By, (Calibration Engineer) M.RAGUL		Authorised by: (Quality Manager/Chief Executive) C.SIVABALAN
--	--	---