



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name **DIVINE TOUCH MEDI CLINIC** Lab No **16585**  
Constituent Group **Chemistry II** Date of Result Entered : **12/04/2023**  
PT item **Lyophilized human serum based** Date of Report Published : **06/05/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	524	149.18	8.49	12.67	149.91 mg/dL	0.06	1.11
2	UREA II	Urease UV / GLDH II	Agape	479	37.53	11.09	4.16	37.89 mg/dL	0.09	0.38
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	394	2.48	9.40	0.23	3.12 mg/dL	2.75	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	480	3.12	16.43	0.51	3.12 mg/dL	0.00	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	371	5.54	8.14	0.45	5.54 g/dL	0.00	0.05
6	ALBUMIN II	BCG - Colorimetric II	Agape	367	3.38	9.46	0.32	3.63 g/dL	0.78	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	449	5.54	14.33	0.79	5.88 mg/dL	0.43	0.07
8	CHOLESTROL II	CHOD-PAP II	Agape	523	117.23	11.15	13.08	110.06 mg/dL	-0.55	1.14
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	386	103.22	12.42	12.82	101.62 mg/dL	-0.12	1.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**DIVINE TOUCH MEDI CLINIC**  
**COLONEL CHOWMUHANI, AGARTALA**  
**TRIPURA**  
**TRIPURA799001**

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*