



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name	Govt. UPHC, Renganathapuram	Lab No	12261
Constituent Group	Chemistry III	Date of Result Entered :	20/07/2023
PT item	Lyophilized human serum based	Date of Report Published :	07/08/2023

Sl. No.	Parameter	Method	Result	Mean	SD	CV	Target	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	1876	109.35	5.55	6.07	102 mg/dL	-1.21	0.28
2	UREA III	Others (Urease Berthelot / Ned Dye) III	47	52.80	7.28	3.85	49 mg/dL	-0.99	1.12
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1850	5.83	6.00	0.35	5.2 mg/dL	-1.80	0.02
4	CHOLESTROL III	CHOD-PAP III	1776	113.00	4.12	4.66	101 mg/dL	-2.58	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Renganathapuram
THE MEDICAL OFFICER
Chengalpattu
TAMILNADU

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Dr. C. Vijayalaxmi
Medical Officer
Urban Primary Health Centre
Renganathapuram,
Tambaram, Municipality