



### CALIBRATION REPORT

Report No: HBMS/CAL/23-24		Calibration Date: 03.10.2023
		Calibration Valid: 02.04.2024
<b>I.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, PIRMARY HEALTH CENTRE, VANAVASI, SALEM.
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	THERMOSCIENTIFIC
II	Model	FINNPIPETTE F3
III	Value	100µl
IV	Serial No	VV/LAB/MP/002

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	100µl
1	100.05
2	100.06
3	100.01
4	99.97
5	100.01
6	100.02

For Heart Beat Medical Systems,



*(Handwritten signature)*

(Authorized signatory)