



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name **DIVINE TOUCH MEDI CLINIC** Lab No **16585**
Constituent Group **Chemistry II** Date of Result Entered : **09/06/2023**
PT item **Lyophilized human serum based** Date of Report Published : **04/07/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	515	271.61	7.90	21.46	262.37 mg/dL	-0.43	1.89
2	UREA II	Urease UV / GLDH II	Agape	461	46.46	10.75	5.00	48.2 mg/dL	0.35	0.47
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	384	0.96	15.15	0.15	1.22 mg/dL	1.78	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	445	4.90	12.87	0.63	5.01 mg/dL	0.17	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	356	5.87	7.51	0.44	5.67 g/dL	-0.45	0.05
6	ALBUMIN II	BCG - Colorimetric II	Agape	358	3.50	8.60	0.30	3.59 g/dL	0.30	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	381	4.42	16.29	0.72	3.97 mg/dL	-0.62	0.07
8	CHOLESTROL II	CHOD-PAP II	Agape	518	124.11	10.85	13.46	111.91 mg/dL	-0.91	1.18
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	382	158.02	10.39	16.42	154.19 mg/dL	-0.23	1.68

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DIVINE TOUCH MEDI CLINIC
COLONEL CHOWMUHANI, AGARTALA
TRIPURA
TRIPURA799001

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****