



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



Lab Name	Govt. Block PHC, Andimadam	Lab No	4600
Constituent Group	Chemistry III	Date of Result Entered :	12/01/2023
PT item	Lyophilized human serum based	Date of Report Published :	09/02/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	148 mg/dL	3.68	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	1.8 mg/dL	-1.85	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	160 mg/dL	7.08	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. Block PHC, Andimadam
BLOCK MEDICAL OFFICER GOVT.COMMUNITY HEALTH CENTRE ARIYALUR DISTRICT
Anadimadam
TAMILNADU621801

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Raj
Milohingia