



# PIPETTE CALIBRATION REPORT

Product code: 4640060  
Description: Finnpiptette F3 100-1000µl  
Pipette SN: RW13708  
Tip: Finntip 1000  
Nr. of channels: 1  
Module SN: -

Manufacturer: Thermo Fisher Scientific, India Pvt. Ltd., D-96, C Road, MIDC Satpur, Nashik - 422007  
Tested by: 10829355

Date:	2/11/20	Time:	3:27:02 PM
Balance:	Mettler XP205	Balance SN:	B115129036, Calibrated 13.11.19
Environment device:	Almemo 2590-4AS	Env dev SN:	H19070544, Calibrated 07.10.20
Air temperature:	26.06 °C	Water temperature:	25.84 °C
Humidity:	60.0 %	Air pressure:	942.6 hPa
Z-factor:	1.0042		

Volume:	100 µl	1000 µl
1	99.6	1001.3
2	99.9	1001.8
3	100.0	1001.1
4	100.2	1001.6
5	100.0	1000.6

Mean	100.0 µl	1001.3 µl
A	0.0 µl	1.3 µl
s	0.2 µl	0.4 µl
Acc%	-0.05	0.13
CV%	0.20	0.04
	<b>PASSED</b>	<b>PASSED</b>

#### Manufacturer specifications:

	Min. volume:	Max. volume:
A±	100 µl	1000 µl
s	1.5 µl	5 µl
Acc%±	0.6 µl	2 µl
CV%	1.50	0.50
	0.60	0.20

Manufacturer specifications meet the requirements of ISO 8655