

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Name: Govt. Block PHC, Andimadam      Lab No: 4600  
 Patient: Chemistry III      Date of Result Entered: 20/12/2022  
 Specimen: Lyophilized Serum      Date of Report Published: 05/01/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1844	239.20	6.25	14.96	239 mg/dL	-0.01	0.70
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1839	3.91	8.24	0.32	3.2 mg/dL	-2.20	0.02
CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1766	114.94	5.41	6.21	123 mg/dL	1.30	0.30

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. Block PHC, Andimadam  
 BLOCK MEDICAL OFFICER GOVT.COMMUNITY HEALTH CENTRE ARIYALUR DISTRICT  
 Andimadam  
 TAMILNADU621801

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

