



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt P.H.C. Ammapattinam
 Consultant Order: Chemistry II
 PT No: Lysophilized human serum based

Lab No: 12518
 Date of Result Entered: 20/06/2023
 Date of Report Published: 04/07/2023

Sl. No	Analyte	Method / Reagent Name	Reference Range	No. of Performances	CV	Performance		Total Score	SD	Z
						CV	SD			
1	GLUCOSE II	GGD-PGD II	Reference (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	264 eqpt/II	-0.54	0.69
2	CREATININE II	Jaffee Kinetic - Alkaline Phosphate II	Reference (Semi automatic / Fully Automatic)	1848	2.97	6.64	8.17	27 eqpt/II	0.76	0.91
3	CHOLESTEROL II	CHOD-PAP II	Reference (Semi automatic / Fully Automatic)	1804	113.73	4.88	5.97	113 eqpt/II	-1.14	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within 1.01 to 2.00	Good
Within 2.01 to 3.00	Accept with caution Warning Signal
Beyond 3.00	Unacceptable performance Action Signal

LAB ADDRESS:
 Govt. P.H.C. Ammapattinam
 Ammapattinam
 Ammapattinam
 TAMILNADU 614617

Coordinator Contact Details:
 Email: cmceqas@cmcvellore.ac.in
 Contact Number: 0416-2282102

Parvati Chandrasekar
 Dr. Parvati Chandrasekar
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed
 Data in CMC EQAS reports is confidential
 CMC EQAS does not make contact any components
 ***** End of Report *****

M. Nargeen Fatima
 MEDICAL OFFICER
 GOVT PRIMARY HEALTH CENTRE
 AMMAPATTINAM-614617