



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Ammapattinam Lab No: 12538
 Constituent Group: Chemistry III Date of Result Entered: 19/07/2023
 PT Item: Lympholized human serum based Date of Report Published: 07/08/2023

Sl. No.	Parameter	Method	Instrument	Result	Unit	CV (%)	SI	SI Unit	SI CV (%)	SI Unit
1	GLUCOSE III	ODD-POD III	Robark Semi-automatic / Fully Automatic	1876	109.35	5.55	6.07	112 mg/dL	0.44	0.28
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robark Semi-automatic / Fully Automatic	1850	5.03	6.00	0.35	4.1 mg/dL	0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robark Semi-automatic / Fully Automatic	1776	113.00	4.12	4.66	118 mg/dL	0.64	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.49	Accept with caution - Warning Signal
Beyond ±2.50	Unacceptable performance - Action Signal

LAB ADDRESS:
 Govt. PHC, Ammapattinam
 Ammapattinam
 Ammapattinam
 TAMILNADU 614617

Coordinator Contact Details:
 Email: etq@cmcvellore.ac.in
 Contact Number: 0436-2282102

Patela Christodoss
 Dr. Patela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed
 Data in CMC EQAS reports is confidential
 CMC EQAS does not wish contract any components
 ***** End of Report *****

M. Nasreen Fatima
 MEDICAL OFFICER
 GOVT PRIMARY HEALTH CENTRE
 AMMAPATTINAM-614617