



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name	Govt. PHC, Aranamattanam	Lab No	12538
Constituent Group	Chemistry III	Date of Result Entered	20/08/2023
PT Item	Lysophilized human serum based	Date of Report Published	01/09/2023

Sl. No.	Parameter	Method	Manufacturer	Observed	Target	Delta	Delta 2	Delta 3	Delta 4	Delta 5	Delta 6
1	GLUCOSE III	DOO-POD III	Roche (Semi-automatic / Fully Automatic)	1841	175.50	5.13	9.00	105 mg/dL	1.16	0.42	
2	CREATININE III	Jaffes Nitrate - Alkaline Picrate II	Roche (Semi-automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.24	0.01	
3	CHOLESTROL III	CHOD-PAP III	Roche (Semi-automatic / Fully Automatic)	1764	113.80	4.07	4.64	114 mg/dL	2.04	0.23	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal
Beyond ±3.0	Unacceptable performance. Action Signal

LAB ADDRESS:  
Govt. PHC, Aranamattanam  
Aranamattanam  
TAMILNADU 614617

Coordinator Contact Details:  
Email: [cmceqas@cmc.ac.in](mailto:cmceqas@cmc.ac.in)  
Contact Number: 0416-2263102

*Parvati Christoforo*  
Dr. Parvati Christoforo  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*M. Naygen Ething*  
MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
ARANAMATTANAM  
MEDICAL OFFICER