

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Vanapuram	Details About Robust Analysis
Lab No	13116	Detail About Monthly Summary
Month	September	Detail about SDI
Year	2022	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 19/09/2022


Date of Report Published : 05/10/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1846	223.49	6.67	14.90	217 mg/dL	-0.44	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1834	3.49	9.05	0.32	3.1 mg/dL	-1.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	112.08	5.53	6.20	212 mg/dL	16.12	0.29

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


MEDICAL OFFICER
 Govt. Primary Health Centre,
 Vanapuram,
 Kallakurichi Dt.