



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - NOVEMBER 2022



PC-102A

Lab Name Govt. PHC, Vanapuram

Lab No 13116

Constituent Group Chemistry III

Date of Result Entered : 18/11/2022

PT Item Lyophilized Serum

Date of Report Published : 06/12/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	82.95	8.62	7.15	89.3 mg/dL	0.89	0.33
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1843	6.98	7.13	0.50	6.5 mg/dL	-0.96	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1767	112.60	5.74	6.46	117.1 mg/dL	0.70	0.31
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Vanapuram
VANAPURAM PHC
Vanapuram
TAMILNADU606205

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Paula Christudas
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Yellore

MEDICAL OFFICER
Govt. Primary Health Centre,
Vanapuram,
Kalthakurichi Dt.

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components