

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13309

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
VELLAYAPURAM, RAMANATHAPURAM DISTRICT.

SRF No.	SRF/23/00717-0003
SRF Date	07-10-2023
Date of Receipt	06-10-2023
Date of Calibration	07-10-2023
Due Date for Calibration	06-10-2024
Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMOSCIENTIFIC
Range	100-1000ul	Model	FINNPIPETTE F3
Resolution	1ul	Material	PVC
Serial Number	RW13134	Operating Range	100-1000ul
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.6	Humidity (%RH)	58	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVC SPL22/12/2115-01	A&D & CH 252	National Standards	19-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		1	2	3	4	5				
1	100	99.85	99.87	99.85	99.84	99.95	99.89	-0.11	0.04	0.47
		99.89	99.95	99.90	99.87	99.90				
2	500	499.90	499.95	499.90	499.90	499.95	499.90	-0.10	0.03	0.47
		499.89	499.88	499.85	499.85	499.90				
3	1000	999.89	999.88	999.80	999.87	999.90	999.91	-0.09	0.03	0.47
		999.93	999.90	999.91	999.94	999.95				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By

Authorised by