

VIEW LAB MONTHLY SUMMARY

Lab Name            Govt. PHC, Kutchanur

Lab No                12706

Month                April

Year                  2023

Constituent Group    Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	112 mg/dL	-1.03	0.34
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	169	42.32	10.68	4.52	44 mg/dL	0.37	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	38	4.58	9.59	0.44	4.4 mg/dL	-0.41	0.14
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	107 mg/dL	-0.80	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

VIEW LAB MONTHLY SUMMARY

Lab Name            Govt. PHC, Kutchanur

Lab No                12706

Month                 August

Year                    2023

Constituent Group    Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	185 mg/dL	1.05	0.42
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	174	28.29	15.20	4.30	37 mg/dL	2.03	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	40	5.43	10.18	0.55	5.2 mg/dL	-0.42	0.17
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

VIEW LAB MONTHLY SUMMARY

Lab Name            Govt. PHC, Kutchanur

Lab No                12706

Month                 July

Year                    2023

Constituent Group    Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	119 mg/dL	1.59	0.28
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	184	51.41	11.14	5.73	48 mg/dL	-0.60	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.6 mg/dL	-0.66	0.02
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	30	5.35	9.89	0.53	5.9 mg/dL	1.04	0.19
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	110 mg/dL	-0.64	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kutchanur	<a href="#">Details About Robust Analysis</a>
Lab No	12706	<a href="#">Detail About Monthly Summary</a>
Month	June	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	282 mg/dL	-0.31	0.69
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	192	32.69	12.96	4.24	31 mg/dL	-0.40	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.2 mg/dL	-2.16	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	34	6.25	4.51	0.28	6.1 mg/dL	-0.53	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	118 mg/dL	0.84	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name            Govt. PHC, Kutchanur  
 Lab No                12706  
 Month                 May  
 Year                    2023  
 Constituent Group    Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 21/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	176 mg/dL	-0.52	0.44
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	163	27.91	13.39	3.74	20.4 mg/dL	-2.01	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.75 mg/dL	-0.08	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	40	4.98	10.94	0.55	4.6 mg/dL	-0.70	0.17
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kutchanur  
Lab No 12706  
Month September  
Year 2023  
Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	310 mg/dL	-0.06	0.69
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	187	35.19	12.73	4.48	29 mg/dL	-1.38	0.66
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	35	6.72	4.36	0.29	6.9 mg/dL	0.61	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	116 mg/dL	0.55	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



Site Maintained by

**Swiftware Solutions**

Best Viewed at pixels 1024 X 768