

			CAL	IBRA	TION	CER	TIFICA	TE	
CERTIFICATE NO: SBS/CL/22/12924					MEDICAL DEVICES				Page No:1 of 1
Issue Date					12-11-2022				
SRF No & Date					SRF/22/00284-0003 & 10-11-2022				
Receipt Date					10-11-2022				
Calibration Date					10-11-2022				
Calibration Due					09-11-2023				
Custo	omer Name & Addre	SS			- 12				
GOV	ERNMENT URBAN	PRIMA	RY HEALTH	CENTRE,					
NELL	IKUPPAM-607105.								
				Details of	Device U	nder Calibra	ation (DUC)		
Description : SEMI AUTO ANALYZER				ZER	: Make & Model		ROBONIK & PR	IETEST TOUCH	
Range :					Sr. N	0	:	AT0890609RBK	
Resolution :					Ident	fication No	:		
DUC Condition : SATISFACTORY				Locat	ion		LABORATORY		
			Enviror	nmental Cor	ditions	& Calibratio	n Procedure	Details	
Environmental Details Temperature: 2			e: 25,3° C	25.3° C Rela		midity	54% RH		
Calibration Procedure No SBS/CP/MD/2			/20	Calibration		done at	one at ONSITE		
				Refe	erence S	tandards De	etails		
S.No	Description	cription Make/ S			te/ SI No:		Certificate No		Validity
1	1 Electrical Safety Analyser Rigel M			Rigel Medi	dical & 44L-1059		TSC/22-23/7400-3		10-08-2023

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	80	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	115	5.9
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500µAfor B,BF,CF	141	11.4

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By Authorised Signatory

S.laus (Calibration Engineer) S.LAKSHMANAN MURUGAN



Quality Manager	Chief Executive
(D.VETRI SELVI)	