



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/22/12936

Page. No: 1 of 1

**Customer Name & Address**

GOVERNMENT URBAN PRIMARY HEALTH CENTRE, NELLIKUPPAM-6007105.	SRF No.	SRF/22/00284 & 0015
	SRF Date	12-11-2022
	Date of Receipt	12-11-2022
	Date of Calibration	12-11-2022
	Due Date for Calibration	11-11-2023
	Issue Date	14-11-2022

**Details of Unit Under Calibration**

Description	Micro Pipette	Make	THERMOSCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1 µl	Material	PVC
Serial Number	RW17040	Operating Range	
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSP121/12/1587-01	A&D & GH-252	National Standards	10-12-2022

Z Factor: 1.00319

Result of Calibration in µl										
SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.86	99.87	99.86	99.87	99.86	99.89	-0.12	0.02	0.47
		99.91	99.90	99.91	99.90	99.91				
2	500	499.96	499.95	499.96	499.95	499.96	499.88	-0.12	0.08	0.47
		499.80	499.81	499.80	499.81	499.80				
3	1000	999.84	999.85	999.84	999.85	999.84	999.84	-0.16	0.01	0.47
		999.83	999.82	999.83	999.82	999.83				

**Remarks**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)  
 P.MYILSAMY



Authorised by:

(Quality Manager/Chief Executive)  
 D.VETRI SELVI