



## CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/22/12922	MECHANICAL	Page No:1 of 1
Issue Date	12-11-2022		
SRF No & Date	SRF/22/00284-0001 & 10-11-2022		
Receipt Date	10-11-2022		
Calibration Date	10-11-2022		
Calibration Due	09-11-2023		

**Customer Name & Address**

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,  
NELLIKUPPAM-607105.

**Details of Device Under Calibration (DUC)**

Description	CENTRIFUGE	Make & Model	B.D. INSTRUMENTATION & BDI-152
Range	3500 RPM	Sr. No	NA
Least Count	5 RPM	Identification No	
DUC Condition	Satisfactory	Location	LABORATORY

**Environmental Conditions & Standard Operating Procedure Details**

Environmental Details	Temperature: 25.3 °C	Relative Humidity	45 % Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	Lutron & DT-2238	JRPM-CCTR-A&S-2022-0036	03-10-2023

**CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY ( ± )
	RPM	RPM	RPM	%
1	100	99.5	-0.5	6
2	2000	1998.4	-1.6	6
3	3000	2998.2	-1.8	6

**REMARKS**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)

S.LAKSHMANAN MURUGAN



Authorised Signatory

Quality Manager  
(D.VETRI SELVI)

Chief Executive