

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Vilanthai  
 Lab No 13113  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	L
					CV	SD			
1	GLUCOSE III	GOD-POD III	1841	175.58	5.13	9.00	195 mg/dL	2.16	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	1764	113.80	4.07	4.64	118 mg/dL	0.91	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Home	My Profile	Submit Lab Result	View Lab Monthly Summary
Lab Summary	Change Password	Home	Sign out

*J. Poornima Subramanian*

**MEDICAL OFFICER**  
**GOVERNMENT PRIMARY HEALTH CENTRE**  
**VILANDAI.**