

9/23, 11:52 PM

Screenshot_20230810-121917.png

July - 2023



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. P.H.C. Govindanagarim Details About Robust Analysis
 Lab No: 12714 Detail About Monthly Summary
 Month: July Detail about SDI
 Year: 2023
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

Date of Result Entered: 15/07/2023
 Date of Report Published: 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Total Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	104.6 mg/dl	-0.45	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	91.41	11.14	5.73	52.35 mg/dl	0.20	0.84
3	CREATININE III	Jaffe Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.7 mg/dl	-0.37	0.02
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	30	5.35	9.89	0.53	4.5 mg/dl	-1.61	0.19
5	CHOLESTROL III	OHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.65	108.8 mg/dl	-0.90	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution, Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Home	Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Forgot Summary	Change Password	Home	Sign out	

Maintained by **Swiftware Solutions** Best Viewed at pixels 1024 x 768

சுருட்டி
 அரசு ஆரம்ப சுகாதார சேவையம்
 கோவிந்தசுரேசுரம், தேனி, தமிழ்நாடு