

			CALIE	BRAT	TION	CERT	IFICAT	<u>E</u>	
CERTIFICATE NO: SBS/CL/23/12886			ME		CHANICAL	Pa	age No:1 of 1		
Issue Date					06-10-2023				
SRF No & Date					SRF/23/00638-0002 & 05-10-2023				
Receipt Date					05-10-2023				
Calibration Date					05-10-2023				
Calibration Due					04-10-2024				
Custor	ner Name & Add	ress							
GOVER	RNMENT PRIMAF	RY HEALTH	CENTRE,						
SILLAN	1ARATHUPATTI-	625521,THE							
			Deta	ails of D	evice Un	der Calibra	tion (DUC)		
Description : CENTRIFUGE				Make & Model		:	LABOTECH & BDI-153		
Range 3500 RPM			Sr. No		:	76529			
Resolution 1 RPM			Identification No		:	NA			
DUC Condition Satisfactory			Location			LABORATORY			
2000	orialion .	Env	ronmental C	onditio	ns & Star	ndard Opera	ating Proced	ure Details	
Environmental Details Temperature: 24									
Calibration Procedure No SBS/CP/ML/0				2		done at ONSITE			
				Refe	rence Sta	andards De	tails		
S.No	No Description			Make/ SI No:		Certificate No		Validity	
1	Digital Tachometer			Lutron & DT-2238		JRPM-CCTR-A&S-2022-0036		03-10-2023	

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.8	0.2	4.2	
2	2000	1998	2	4.2	
3	3000	2998	2	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

(Calibration Engineer) P.PRASANNA

Calibrated By



(C.SIVABALAN)

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS