

## VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. UPHC, Masthanpatti**[Details About Robust Analysis](#)Lab No **10795**[Detail About Monthly Summary](#)Month **March**[Detail about SDI](#)Year **2023**Constituent Group **Chemistry III**[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)Date of Result Entered : **20/03/2023**Date of Report Published : **04/04/2023**

| Sl.No | Analyte                        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-------|--------------------------------|---------------------------------------|----------------------------------------------|--------------------|--------|--------------|------|------------|-------|------|
|       |                                |                                       |                                              |                    |        | CV           | SD   |            |       |      |
| 1     | <a href="#">GLUCOSE III</a>    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1878               | 91.48  | 7.46         | 6.82 | 99 mg/dL   | 1.10  | 0.31 |
| 2     | <a href="#">CREATININE III</a> | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1851               | 1.33   | 9.08         | 0.12 | 1.3 mg/dL  | -0.25 | 0.01 |
| 3     | <a href="#">CHOLESTROL III</a> | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1761               | 110.93 | 5.23         | 5.80 | 108 mg/dL  | -0.50 | 0.28 |

| SDI Range                       | Interpretation                           |
|---------------------------------|------------------------------------------|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)

---

[Yearly Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)

---