

VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. UPHC, Masthanpatti**[Details About Robust Analysis](#)Lab No **10795**[Detail About Monthly Summary](#)Month **April**[Detail about SDI](#)Year **2023**Constituent Group **Chemistry III**[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)Date of Result Entered : **19/04/2023**Date of Report Published : **06/05/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	128 mg/dL	1.19	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	107 mg/dL	-0.80	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)

[Yearly Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)
