

VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. UPHC, Masthanpatti**[Details About Robust Analysis](#)Lab No **10795**[Detail About Monthly Summary](#)Month **May**[Detail about SDI](#)Year **2023**Constituent Group **Chemistry III**[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)Date of Result Entered : **20/05/2023**Date of Report Published : **05/06/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	178 mg/dL	-0.31	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	117 mg/dL	0.41	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)

[Yearly Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)
