

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Masthanpatti

[Details About Robust Analysis](#)

Lab No 10795

[Detail About Monthly Summary](#)

Month July

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.1 mg/dL	0.77	0.02
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	120 mg/dL	1.50	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)

---

[Yearly Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)

---