

VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. Block PHC, Gurusvareddiyur	<b>Details About Robust Analysis</b>
<b>Lab No</b>	7375	<b>Detail About Monthly Summary</b>
<b>Month</b>	August	<b>Detail about SDI</b>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	


Click on the analyte to view Graphical Data

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	173 mg/dL	-0.29	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

  
**Medical Officer**  
 Primary Health Center,  
 Gurusvareddiyur,  
 Erode - 639 111.

12-10-20