
| CERTIFICATE | O: SBS | S/CL/23/13893 | | | MECHANICAL | Danie | Mad of a |
|--|-----------------|------------------------|----------------------|---|----------------|----------------------|----------|
| Issue Date SRF No & Date Receipt Date Calibration Date | | | SRF/2: 09-10-2 | MECHANICAL Page No:1 of 1 09-10-2023 SRF/23/00834-0001 & 09-10-2023 09-10-2023 09-10-2023 | | | |
| Calibration Due | | | | 08-10-2024 | | | |
| Customer Name | | | And the second | | | | |
| GOVERNMENT P NEERMULAI-6147 | | | ICT. | | | | |
| | | | Details of Device U | nder Calibra | rtion (DUC) | | |
| Description : CENTRIFUGE | | Make 8 | Make & Model : | | LABOMED & NA | | |
| SOUGHT IN | | | Sr. No | | : | NA | |
| esolution : 1 RPM | | | Identific | Identification No | | NA | |
| UC Condition : Satisfactory | | | Locatio | | | LABORATORY | |
| | E | | Conditions & Sta | ndard Oper | rating Procedu | ure Details | |
| rvironmental Details Temperatu | | re: 25.3°C Relativ | | umidity | 54% Rh | | |
| libration Procedure No SBS/CP/MI | | J04 Calibration | | done at ONSITE | | | |
| _ | The same of the | | Reference St | andards De | tails | | |
| o Description | | Make/ SI No: | | Certificate No | | Validity | |
| Digital Tachometer | | LINE SEIKI / 175-0034V | | JRPM-CCTR-A&S-2023-0013 | | 09-06-2024 | |
| BRATION RESU | | | | | | | 1 |
| DEVICE UNDER CALIBRATION S | | STANDAR | STANDARD INSTRUMENTS | | VIATION | EXPANDED UNCERTAINTY | |
| RPM | | | RPM | | RPM | % | |
| 1000 | | | 999.3 | | 0.7 | 4.2 | |
| 2000 | | | 1999.4 | | 0.6 | 4.2 | |
| 3000 | | | 2998.7 | | 1.3 | 4.2 | |
| RKS | | | | | | | 7.2 |

- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor

Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY



(C.SIVABALAN)

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS