

CALIBRATION CERTIFICATE


Certificate Number :	MED/R/177	Date of issue:	29-Sep-23
ULR No.	CC35352300000177F	Job Order/ CRF No. :	
Calibration Date	28-Sep-23	Calibration Due	27-Sep-24

CUSTOMER INFORMATION		UNIT UNDER CALIBRATION	
Name	Upgraded Govt Primary Health centre	Description	CENTRIFUGE
Address	Guruvareddiyur Erode	Range / Resolution	1000-6000 RPM
		Make	BD INSTRUMENTATION
		Model	BDI-152
		Serial No.	20111304
		Identification No	UGPHC/LAB/CENT/01
		Location	LAB


ENVIRONMENTAL CONDITION		MASTER EQUIPMENT DETAILS	
Temperature	27 ± 2°C	Description	DIGITAL TACHOMETER
Humidity	50 ± 10 % Rh	Certificate No	CC33882300000124F
Air Pressure	1013 mbar	Identification No	S320211
		Calibration Due	18-Aug-24
		Traceability	National Standards

SLNo	UUC READING (RPM)	STANDARD READING (RPM)	Error (RPM)	Expanded Uncertainty ± RPM
1	1000	1003	3	3.00
2	2000	2006	6	3.00
3	3000	2996	4	3.00
4	4000	3996	4	3.00
5	6000	5992	8	3.00

- Note:**
- The UUC is calibrated at Lab.
 - This calibration certificate relates only to the UUC supplied.
 - Calibration certificate shall not be reproduced, except in full, without written approval of medex labtech solutions.
 - Results reported are valid at the time and under stated environmental conditions of calibration.
 - The uncertainty of measurement is expressed at 95.45% confidence level with coverage factor k=2
 - This calibration certificate is issued for scientific or industrial purpose only.
 - Unless Otherwise Specified the Measurement Data Reported is "As Found" -without any adjustment.
 - Next calibration due as per customer's plan and requirement.
 - Decision Rule Required : No

Calibrated by

 Vipin PR
 Calibration Engineer



Authorized By

 Jude Joseph Raj
 Quality Manager

X/657/28, 2nd Floor, RP Tower, Opp. Chaldean Centre, Pallikulam Road, Thrissur, Kerala, India - 680001
 Email: sales@medexlabtech.com, service@medexlabtech.com
 Mob: 9949037588562, 8113816000