

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Name:	M/S.Government Urban Primary Health Centre, 5/307, Kabeer Nagar, Masthanpatti Road	Certificate No:	AMS/HOSP/289/0991
Address:	Masthanpatti, Madurai - 625 020.	Issuing Date:	28/04/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

Device:	Centrifuge	Id. No:	PHCMP/LAB/CEN/001
Make:	REMI	Department:	Laboratory
Model:	R.8C	Calibrated On:	27/04/2022
Sl. No:	ZDEN - 12168	Calibration Due Suggested:	26/04/2023
Type:	Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 4500RPM		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/22/ATL/0432-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023
3	Stop Watch	Nivia / JS 307	PC/01/22/ATL/0134-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	1390	1388	-2.0
2	1500	1505	+5.0
3	2000	1997	-3.0
4	2940	2937	-3.0
5	3560	3562	+2.0

Allowable Tolerance: Set by Customer

Time: (Minutes)

S.No	DUC Reading (Minutes)	Standard Reading (Minutes)	Correction(Seconds)
1	5	5'10"	+10"
2	6	6'11"	+11"
3	8	8'11"	+11"
4	9	9'11"	+11"
5	10	10'10"	+10"

Allowable Tolerance: Set by User

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. S. S.

aceagm@gmail.com



www.acemedical.in

Approved by:
(Technical Manager)

B. S. S.

93455 56140, 93455 56148