

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Podakudi

Lab No 12844

Month August

Year 2023

Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SO			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	213 mg/dL	4.16	0.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	22 mg/dL	-1.46	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

B.1028
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
PODAKKUDI-D.P. Code: 380K
THIRUVARUR HUD.

10:38 PM

LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Podakudi
 Lab No: 12644
 Month: August
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	213 mg/dL	4.16	0.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	22 mg/dL	-1.46	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.84	110 mg/dL	-0.82	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

B. [Signature]
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
PODAKKUDI-OP Code: 380K
THIRUVARUR NUD.

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Podakudi
 Lab No 12844
 Month July
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 15/07/2023


Date of Report Published : 07/08/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	109 mg/dL	-0.06	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	51.41	11.14	5.73	42 mg/dL	-1.64	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.7 mg/dL	-0.37	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	122 mg/dL	1.93	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 PODAKKUDI-DF, Code: 322
 THIRUVARUR HSD.

MONTHLY SUMMARY REPORT

Lab Name	Lab No. / Address
Lab No.	Lab
Lab Name	Lab
Lab	Lab
Lab	Lab
Lab	Lab

The following applies to the following data

Lab Name	Lab No.	Address - Reporting Lab	Reporting Period		No. of Test Results		Passing		Total Value		%	
			Start Date	End Date	Pass	Fail	Pass	Fail	Pass	Fail		
Lab Name A	Lab No. A	Address A	2023-01-01	2023-01-31	100	0	100	0	100	0	100	0
Lab Name B	Lab No. B	Address B	2023-01-01	2023-01-31	50	50	50	50	50	50	50	50
Lab Name C	Lab No. C	Address C	2023-01-01	2023-01-31	20	80	20	80	20	80	20	80
Lab Name D	Lab No. D	Address D	2023-01-01	2023-01-31	10	90	10	90	10	90	10	90
Lab Name E	Lab No. E	Address E	2023-01-01	2023-01-31	5	95	5	95	5	95	5	95
Lab Name F	Lab No. F	Address F	2023-01-01	2023-01-31	2	98	2	98	2	98	2	98
Lab Name G	Lab No. G	Address G	2023-01-01	2023-01-31	1	99	1	99	1	99	1	99
Lab Name H	Lab No. H	Address H	2023-01-01	2023-01-31	0	100	0	100	0	100	0	100
Lab Name I	Lab No. I	Address I	2023-01-01	2023-01-31	0	100	0	100	0	100	0	100
Lab Name J	Lab No. J	Address J	2023-01-01	2023-01-31	0	100	0	100	0	100	0	100

Health Services: _____
 Lab Name: _____
 Address: _____
 City: _____
 State: _____
 Zip: _____
 Contact: _____
 Date: _____

MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 PODAKKUDI-09, C-58, 180A
 THIRUVARUR HSD.

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Podakudi
 Lab No 12844
 Month April
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	126 mg/dL	0.91	0.34
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	34 mg/dL	-1.84	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	104 mg/dL	-1.31	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Early Summary	Change Password	Home	Sign-out

B. Claf
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 PODAKKUDI-DP. Code: 605 011
 THIRUVARUR HSD.