

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Villapuram	<a href="#">Details About Robust Analysis</a>
Lab No	10818	<a href="#">Detail About Monthly Summary</a>
Month	September	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)
[All Analyser Result](#)
[Print](#)

Date of Result Entered : 19/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Transasia / Erba	18	307.20	6.79	20.85	295 mg/dL	-0.59	9.83
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.26	9.52	0.12	1.5 mg/dL	2.00	0.05
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Transasia / Erba	19	113.64	2.93	3.33	114 mg/dL	0.11	1.53

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>