

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Villapuram	Details About Robust Analysis
Lab No	10818	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)
[All Analyser Result](#)
[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	23	173.17	4.83	8.36	180 mg/dL	0.82	3.49
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.74	6.34	0.11	1.8 mg/dL	0.55	0.05
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	17	111.42	3.18	3.54	113 mg/dL	0.45	1.72

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out