

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Villapuram	<a href="#">Details About Robust Analysis</a>								
Lab No	10818	<a href="#">Detail About Monthly Summary</a>								
Month	July	<a href="#">Detail about SDI</a>								
Year	2023									
Constituent Group	Chemistry III									
		<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 20/07/2023								
		Date of Report Published : 07/08/2023								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Transasia / Erba	24	109.83	6.83	7.50	110 mg/dL	0.02	3.06
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	18	5.79	3.00	0.17	6.1 mg/dL	1.78	0.08
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Transasia / Erba	19	112.64	4.38	4.93	115 mg/dL	0.48	2.26
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>