



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manakudayan
 Lab No 12013
 Month August
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 22/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51	195 mg/dL	2.17	2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13	1.8 mg/dL	0.32	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	118 mg/dL	0.81	0.03

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

R-E
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 மணக்குடையான்-621711
 செந்துறை (வ) அரியலூர் (மாவட்டம்)

2023/10/12 16:22

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manakudayan
 Lab No 12013
 Month July
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95	99 mg/dL	-1.80	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27	6 mg/dL	0.41	0.07
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88	120 mg/dL	1.48	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

R. J.
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 மணக்குடையான்-621719
 செந்துறை (வ) அரியலூர் (மாவட்டம்)

2023/10/12 16:22

VIEW LAB MONTHLY SUMMARY

Govt. PHC, Manakudayan

Details About Robust Analysis

12013

Detail About Monthly Summary

June

Detail about SDI

2023

Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
SE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	65	287.55	4.50	12.95	296 mg/dL	0.65	3.21
NINE	Haffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	67	2.58	6.39	0.16	118 mg/dL	721.38	0.04
STROL	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	2.7 mg/dL	-24.12	0.76

DI Range	Interpretation
0 to +1.00	Excellent.
01 to +2.00	Good.
01 to +2.99	Accept with caution. Warning Signal.
0	Unacceptable performance. Action Signal.

2023/10/12 16:23

MONTHLY SUMMARY REPORT - MAY 2023

PC-1024

Lab Name Govt. PHC, Manakudayan

Lab No

12013

Constituent Group Chemistry III

Date of Result Entered :

19/05/2023

PT item Lyophilized human serum based

Date of Report Published :

05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52	188 mg/dL	0.65	2.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12	1.8 mg/dL	0.41	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	119 mg/dL	0.71	0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Manakudayan

THE MEDICAL OFFICER, SENDURAI

R. J.
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 மணக்குடையான்-621719
 செந்துறை (வ) அரியலூர் (மாவட்டம்)

2023/10/12 16:23

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manakudayan
 Lab No 12013
 Month April
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	74	118.33	5.77	6.83	116 mg/dL	-0.34	1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	65	2.14	7.02	0.15	118 mg/dL	772.4	0.04
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79	2.1 mg/dL	-22.78	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address

Modify Methods

Submit Lab Results

View Lab Monthly Summary

R.I.
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 மணக்குடையான்-621719
 செஞ்சுறை (வ) அரியலூர் (மா.வ)

2023/10/12 16:23

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manakudayan
 Lab No 12013
 Month March
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	74	92.52	7.71	7.13	90 mg/dL	-0.35	1.66
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	70	1.32	7.52	0.10	1.4 mg/dL	0.81	0.02
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	111.55	4.29	4.78	118 mg/dL	1.35	0.79

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

R.I.
 மருத்துவ அலுவலர்
 ஆக் ஆரம்ப சுகாதார நிலையம்
 மணக்குடையான்-621719
 மருத்துரை (வ) அரியலூர் (கா.வ)

2023/10/12 16:23

Links

Modify Address

Modify Methods

Submit Lab Results

View Lab Monthly Summary