

APRIL 2023

NO	Test Name	Method	Temp	Lot No
	Glucose	Grid POD	Control	047206
	Creatinine	J.K. Alkaline	EP700	501204
	Cholesterol	CHOD-PAP	EP700	502202

Kit Exp. Date	Date of Test	Result	LT sign	MO sign
10/2023	17-4-23	122.9	ESN	RS
06/2024	17-4-23	28	ESN	RS
01/2024	17-4-23	128.6	ESN	RS

S. No	Reported Result	SDI value	Remark
1	615/03	0.48	Excellen
2	615/03	3.56	un accept
3	615/23	2.95	Accept

RS

Signatures  
 Signatures: average for results  
 Signature: Signature

MAY 2023

Test Name	Method	Temp	Lot No
Glucose	Grid POD	Control	047206

Kit Exp. Date	Date of Test	Result	LT sign	MO sign
10/24	22-5-23	167.8	ESN	RS