


CALIBRATION / TESTING REPORT				
<b>CUSTOMER DETAILS :</b> GOVERNMENT PRIMARY HEALTH CENTRE, SHANMUGANATHAPURAM, SIVAGANGAI.		<b>CERTIFICATE NO</b>	<b>TICS/2023/STM/SI/002</b>	
		<b>EQUIPMENT NAME</b>	<b>SEMI AUTO ANALYZER</b>	
		<b>MANUFACTURE</b>	<b>ROBNIK</b>	
		<b>MODEL</b>	<b>PRIETEST TOUCH</b>	
<b>CALIBRATION DATE</b>	<b>20/09/2023</b>	<b>ASSET NO</b>	<b>ATCD0580321RBK</b>	
<b>CALIBRATION DUE DATE</b>	<b>19/09/2024</b>	<b>INSTALLED</b>	<b>LAB</b>	
		<b>DEPARTMENT</b>		
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No				
<b>TEST EQUIPMENT</b>	<b>MANUFACTURE</b>	<b>MODEL</b>	<b>SERIAL NUMBER</b>	
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011	
<b>PARAMETER</b>	<b>UNITS</b>	<b>VALUE</b>	<b>LIMITS</b>	<b>STATUS</b>
Main Voltage L1-L2	V	230.2	240±10	PASS
Main Voltage L1-G(PE)	V	232.1	240±10	PASS
Main Voltage L2-G(PE)	V	0.9	<5	PASS
Load Current	A	0.321	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.17	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS
Note : Test Standard: IEC 62353 Defibrillator paddles 100 μA, per IEC 60601-2-4				
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.				
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE				
Calibrated By : 		Authorized Sign With Seal 