



| CALIBRATION / TESTING REPORT | | | | | |
|--|-------------------|--|-----------------|----------------------------|-------------------------------------|
| CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE SHANMUGANATHAPURAM, SIVAGANGAI. | | CERTIFICATE NO | | TICS/2023/STM/SI/04 | |
| | | EQUIPMENT NAME | | MICROPIPETTE | |
| | | MANUFACTURE | | FINNPIPETTE | |
| | | MODEL | | 10-100 µL | |
| CALIBRATION DATE | 20/09/2023 | ASSET NO | | STMPH 004 | |
| CALIBRATION DUE DATE | 19/09/2024 | INSTALLED DEPARTMENT | | LAB | |
| 1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No | | | | | |
| TEST EQUIPMENT | | ID. NO | | CERTIFICATE NUMBER | |
| ELECTRONIC SEMI MICRO BALANCE | | NCS/WB01 | | TSC/23-24/5424-1 | |
| TEST REPORT | | | | | Temperature 25°C Humidity RH 55% |
| PARAMETERS | UNITS | SET VALUE | MEASURED VALUES | TOLERANCE | STATUS |
| VOLUME | µL | 50 | 50.21 | ± 1 | PASS |
| | | 100 | 999.16 | ± 1 | PASS |
| *The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards. | | | | | |
| <input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE | | | | | |
| Calibrated By : | | Authorized Sign With Seal | | | |
|  | |  | | | |