

CALIBRATION / TESTING REPORT					
<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE SHANMUGANATHAPURAM, SIVAGANGAI.		<b>CERTIFICATE NO</b>		<b>TICS/2023/STM/SI/05</b>	
		<b>EQUIPMENT NAME</b>		<b>MICROPIPETTE</b>	
		<b>MANUFACTURE</b>		<b>PINNIPIPETTE</b>	
		<b>MODEL</b>		<b>100-1000µL</b>	
<b>CALIBRATION DATE</b>	<b>20/09/2023</b>	<b>ASSET NO</b>		<b>STMPH005</b>	
<b>CALIBRATION DUE DATE</b>	<b>19/09/2024</b>	<b>INSTALLED DEPARTMENT</b>		<b>LAB</b>	
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
<b>TEST EQUIPMENT</b>		<b>ID. NO</b>		<b>CERTIFICATE NUMBER</b>	
<b>ELECTRONIC SEMI MICRO BALANCE</b>		<b>NCS/WB01</b>		<b>TSC/23-24/5424-1</b>	
<b>TEST REPORT</b>					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
<b>VOLUME</b>	µL	100	99.36	± 1	PASS
		1000	999.42	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					
					