

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt.phc kamarajapuram									
Lab No	11564	Back to Search Page								
Month	August									
Year	2023	<input type="button" value="Print"/>								
Constituent Group	Chemistry III									
Date of Result Entered : 19/08/2023										
Date of Report Published : 01/09/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48			3.25
			Transasia / Erba	23	173.17	4.83	8.36			3.49
			Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	180 mg/dL	0.49	0.42
			Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51			2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11			0.04
			Transasia / Erba	21	1.74	6.34	0.11			0.05
			Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
			Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13			0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49			1.56
			Transasia / Erba	17	111.42	3.18	3.54			1.72
			Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	116 mg/dL	0.47	0.22
			Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97			0.81
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

