

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt.phc kamarajapuram									
Lab No	11564	Back to Search Page								
Month	March									
Year	2023	<input type="button" value="Print"/>								
Constituent Group	Chemistry III									
Date of Result Entered : 19/03/2023										
Date of Report Published : 04/04/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84			2.61
			Transasia / Erba	18	92.24	11.77	10.85			5.12
			Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	99 mg/dL	1.10	0.31
			Any Analyser (Automation / Semi Automation)	74	92.52	7.71	7.13			1.66
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13			0.04
			Transasia / Erba	16	1.36	9.29	0.13			0.06
			Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.36 mg/dL	0.25	0.01
			Any Analyser (Automation / Semi Automation)	70	1.32	7.52	0.10			0.02
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68			2.20
			Transasia / Erba	19	111.05	6.05	6.72			3.08
			Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	109 mg/dL	-0.33	0.28
			Any Analyser (Automation / Semi Automation)	148	111.55	4.29	4.78			0.79
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

