



VIEW LAB MONTHLY SUMMARY

Lab Name: **Dont. PHC, Durankudi** [Details About Robust Analysis](#)
 Lab No: **10000** [Detail About Monthly Summary](#)
 Month: **August** [Detail about SDI](#)
 Year: **2023**
 Constituent Group: **Christy III**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 18/08/2023
 Date of Report Published: 01/09/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robosnik (Semi automatic / Fully Automatic)	1841	175.38	2.13	9.00	186 mg/dL	-1.06	0.42
2	CREATININE II	Jaffes Kinetic - Alkaline Picrate II	Robosnik (Semi automatic / Fully Automatic)	1852	1.76	0.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL II	CHOD-PAP II	Robosnik (Semi automatic / Fully Automatic)	1794	113.80	4.07	4.94	115 mg/dL	0.26	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign out



VIEW LAB MONTHLY SUMMARY

Lab Name: **Dist. PhC, Durankudi** [Details About Robust Analysis](#)
 Lab No: **10000** [Detail About Monthly Summary](#)
 Month: **September** [Detail about SDI](#)
 Year: **2023**
 Constituent Group: **Christy III**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 20/09/2023
 Date of Report Published: 09/10/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robosik (Semi automatic / Fully Automatic)	1870	310.80	4.82	14.98	288 mg/dL	-1.00	0.09
2	CREATININE III	Jaffe's Kinetic - Alkaline Picrate II	Robosik (Semi automatic / Fully Automatic)	1852	1.29	0.14	0.10	1.2 mg/dL	-0.98	0.00
3	CHOLESTROL III	CHOD-PAP III	Robosik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	106 mg/dL	-1.77	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Logout



VIEW LAB MONTHLY SUMMARY

Lab Name: **Dont. PHC, Durankudi** [Details About Subst Analysis](#)
 Lab No: **10000** [Detail About Monthly Summary](#)
 Month: **July** [Detail about SDI](#)
 Year: **2023**
 Constituent Group: **Christy III**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 20/07/2023
 Date of Report Published: 07/08/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Rebornik (Semi automatic / Fully Automatic)	1870	109.35	3.53	0.07	108 mg/dL	-0.22	0.28
2	CREATININE II	Jaffe's Kinetic - Alkaline Picrate II	Rebornik (Semi automatic / Fully Automatic)	1850	5.92	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL II	CHOD-PAP II	Rebornik (Semi automatic / Fully Automatic)	1778	113.00	4.12	4.99	116 mg/dL	0.64	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Subst Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign out

Site Maintained by
Swiftware Solutions

Best Viewed at pixels 1024 X 768



VIEW LAB MONTHLY SUMMARY

Lab Name: **Dept. PHC, Surankudi** [Details About Robust Analysis](#)
 Lab No: **10000** [Detail About Monthly Summary](#)
 Month: **June** [Detail about SDI](#)
 Year: **2023**
 Constituent Group: **Chemistry III**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 20/06/2023
 Date of Report Published: 04/07/2023

Sl. No.	Analyte	Method/ Principle Name	Analyser Name	No. of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Roborvik (Semi automatic / Fully Automatic)	1823	2.50	3.17	14.82	290 mg/dL	0.82	0.89
2	CREATININE III	Jaffe's Kinetic - Alkaline Picrate III	Roborvik (Semi automatic / Fully Automatic)	1848	2.57	0.54	0.17	2.8 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Roborvik (Semi automatic / Fully Automatic)	1804	113.72	4.45	5.07	118 mg/dL	0.84	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign out



VIEW LAB MONTHLY SUMMARY

Lab Name: **Dist. PHC, Durankudi** [Details About Robust Analysis](#)
 Lab No: **10000** [Detail About Monthly Summary](#)
 Month: **May** [Detail about SDI](#)
 Year: **2023**
 Constituent Group: **Christy III**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 20/05/2023
 Date of Report Published: 05/06/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robosnk (Semi automatic / Fully Automatic)	1834	180.88	2.19	9.39	183 mg/dl	0.23	0.44
2	CREATININE II	Jaffes Kinetic - Alkaline Picrate II	Robosnk (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.6 mg/dl	0.30	0.01
3	CHOLESTROL II	CHOD-PAP II	Robosnk (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	107 mg/dl	-1.30	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign out



VIEW LAB MONTHLY SUMMARY

Lab Name: Dmt. PKC, Durankudi [Details About Subst Analysis](#)

Lab No: 10000 [Detail About Monthly Summary](#)

Month: April [Detail about SDI](#)

Year: 2023

Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 20/04/2023

Date of Report Published: 06/05/2023

Sl. No.	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Rebornik (Semi automatic / Fully Automatic)	1832	115.45	0.04	7.21	121 mg/dl	0.21	0.34
2	CREATININE III	Jaffe's Kinetic - Alkaline Picrate III	Rebornik (Semi automatic / Fully Automatic)	1820	2.16	0.09	0.18	2.2 mg/dl	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Rebornik (Semi automatic / Fully Automatic)	1750	111.61	5.15	5.79	110 mg/dl	-0.28	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Subst Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Logout