

CENTRIFUGE Calibration / Verification Record Sheet				
Location: ICTC- (Non ANC) CHC - Peñdapuram, KAKINADA DISTRICT		Range : Up to 3500RPM		
Equipment ID: Remi ;		UID .NO :CHC-PDP/ICTC/CEN 1of 1		
Calibration Procedure Reference: In-house verification of centrifuge using NABL calibrated Tachometer				
Date of Calibration: 20.9.2023		Due date of Calibration: 19.9.2024		
Certified centrifuge		Tachometer ID : EAPL ;DT - 2001B Calibration date : 7.6.2023 Calibration date due: 6.6.2024		
Verification Done by: S. Suneela Raja Kumari, Technical Officer, SRL, KMC, Kakinada				
S.NO	STANDARD READING (RPM)	UUC READING (RPM)	ERROR (RPM)	REMARK
1.	1700	1726	26	Acceptable
2.	2000	2052	52	Acceptable
3.	2300	2341	41	Acceptable
4	2700	2732	32	Acceptable
5.	3000	2959	-41	Acceptable
Accept/ Reject: Accepted since average difference is within +/- 100RPM				
Final Interpretation: Centrifuge is in good working condition				
Action Taken: - NA				
Remark: - NA				

M. S. S. S.