




CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - SEPTEMBER 2023



Lab Name: Govt. PHC, Vizhapallam
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 12009
 Date of Result Entered: 20/09/2023
 Date of Report Published: 05/10/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	317 mg/dL	0.41	0.69
2	CREATININE III	Jeffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	111 mg/dL	-0.61	0.20
SDI Range										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Vizhapallam
 THE MEDICAL OFFICER, vizhapallam
 Vizhapallam
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

Dr. J
MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE,
 VIZHAPALLAM - 621 803
 ARYALUR DIST.

Dr. J