


CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS :		CERTIFICATE NO		TICS/2023/MM/SG/PN/02	
GOVT.PRIMARY HEALTH CENTRE		EQUIPMENT NAME		CENTRIFUGE	
PAGANERI,		MANUFACTURE		LABOTECH	
SIVAGANGAL		MODEL		-	
CALIBRATION DATE	03/10/2023	ASSET NO	PN PH 002		
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB		
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		MANUFACTURE		MODEL	
DIGITAL TACHOMETER		KUSAM MECO		KM2234 BL	
ELECTRICAL SAFTY ANALYZER		DATREND		Y Pad-mini	
				S1049875	
				MPM 18040011	
TEST REPORT					
				Temperature 25°C Humidity RH 55%	
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 2	2117	± 5	PASS
		Speed 4	2865		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE Calibrated By:  Authorized Sign With Seal 