


CALIBRATION / TESTING REPORT					
<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE PAGANERI, SIVAGANGAL.		CERTIFICATE NO	TICS/2023/MM/SG/PN/03		
		EQUIPMENT NAME	MICROPIPETTE		
		MANUFACTURE	MICROLUX		
		MODEL	100-1000µL		
CALIBRATION DATE	03/10/2023	ASSET NO	PN PH 003		
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB		
<p>1. Have more dust in machine outside      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required)      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	99.62	± 1	PASS
		1000	999.37	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE Calibrated By:  Authorized Sign With Seal 