

CALIBRATION / TESTING REPORT					
<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE PAGANERI, SIVAGANGAI.		<b>CERTIFICATE NO</b>	TICS/2023/MM/SG/PN/04		
		<b>EQUIPMENT NAME</b>	MICROPIPETTE		
		<b>MANUFACTURE</b>	VISION PLUS		
		<b>MODEL</b>	100 µL		
<b>CALIBRATION DATE</b>	03/10/2023	<b>SERIAL NO</b>	PN LAB 004		
<b>CALIBRATION DUE DATE</b>	02/10/2024	<b>INSTALLED DEPARTMENT</b>	LAB		
<p>1. Have more dust in machine outside      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required)      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p>					
<b>TEST EQUIPMENT</b>	<b>ID. NO</b>	<b>CERTIFICATE NUMBER</b>	<b>VALIDITY</b>		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	99.94	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE Calibrated By: <i>Pr</i> Authorized Sign With Seal					
					