

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Ansari Nagar	<a href="#">Details About Robust Analysis</a>
Lab No	10778	<a href="#">Detail About Monthly Summary</a>
Month	May	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	
<a href="#">Click on the analyte to view Graphical Data</a>		<input type="button" value="All Analyser Result"/> <input type="button" value="Print"/>
Date of Result Entered : 18/05/2023		
Date of Report Published : 05/06/2023		

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	203 mg/dL	2.35	0.44
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.6 mg/dL	-1.21	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	103 mg/dL	-1.99	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>